

Aquarius-Vet

Veterinary Monitor



Committed to Excellence

Configuration

Optional

SpO2 + NIBP, Li-ion battery

Masimo/Nellcor SpO2, Quick Temp, Bar code scanner, wired/wireless CMS

SpO2+NIBP+ECG+TEMP, Li-ion battery

Masimo/Nellcor SpO2, EtCO2, Quick Temp, Bar code scanner, Thermal Recorder, wired/wireless CMS



Touch Screen (Optional)



Quick Temp (Infrared Ear Thermometer)



Portable Design



Aquarius-Vet

Veterinary Monitor



- 8" color TFT LCD Screen (Touch screen is an optional)
- Portable, Lighter weight and sturdy design
- Flexible parameters configuration for different clinical environments
- Rechargeable Li-ion Battery (up to 12 hours uninterrupted work)
- Big font and color font display setting
- Spot-check and continuous monitoring mode

- Selectable for Dog, Cat and other animals
- Wired/Wireless CMS, support HL7 protocol to HIS
- Barcode scanner support
- Thermal recorder support
- Graphical & tabular trend review
- 48h full disclosure wave review for each animal (stroed in SD card)

Specifications

Display

8" color TFT LCD Screen, resolution: 600 x 800

ECG

Lead type

- 3-lead: I, II, III
- 5-lead: I, II, III, aVR, aVL, aVF, V

Display sensitivity:

- 2.5mm/mV (x0.25), 5mm/mV (x0.5), 10mm/mV (x1.0), 20mm/mV (x2.0)

Wave sweep speed: 6.25mm/s, 12.5 mm/s, 25 mm/s, 50 mm/s

Bandwidth

- Diagnostic mode: 0.05Hz~100Hz
- Monitor mode: 0.5Hz~40Hz
- Surgery mode: 1Hz~20Hz
- Strong filter mode: 5Hz~20Hz

CMRR: >100dB

Notch: 50/60Hz notch filter can be set to on or off

Differential input impedance: >5MΩ

Electrode polarization voltage range: ±400mV

Baseline recovery time <3s after defibrillation (in monitor and surgery mode)

Calibration signal: 1mV (peak - peak), accuracy ±3%

RESP

Measurement method: Thoracic electrical bioimpedance

Measuring lead: Lead I, II

Wave gain: x0.25, x0.5, x1, x2

Respiratory impedance range: 0.5-5Ω

Baseline impedance: 500-4000Ω

Gain: 10 grades

Scan speed: 6.25mm/s, 12.5 mm/s, 25mm/s

TEMP

Measurement method: Thermistor

Measuring range: 5~50°C (41~122°F)

Resolution: 0.1°C

Measurement accuracy: ±0.1°C

Recorder (optional)

Built-in Thermal dot array

Horizontal resolution: 16 dots/mm (25 mm/s paper speed)

Vertical resolution: 8 dots/mm

Paper speed: 25 mm/s, 50 mm/s

Number of waveform channels: 3

NIBP

Measurement method: Automatic oscillometric method

Operating mode: Manual, automatic, continuous

Measurement unit: mmHg/kPa selectable

Typical measurement time: 20~40s

Measurement type: Systolic, Diastolic, Mean

Measurement range (mmHg)

- | | | |
|-----------------------------|--------|--------|
| Range of Systolic pressure: | Dog | 40-270 |
| | Cat | 40-200 |
| | Others | 40-135 |

- | | | |
|------------------------------|--------|--------|
| Range of Diastolic pressure: | Dog | 10-210 |
| | Cat | 10-150 |
| | Others | 10-95 |

- | | | |
|-------------------------|--------|--------|
| Range of Mean pressure: | Dog | 20-230 |
| | Cat | 20-165 |
| | Others | 20-105 |

Measurement accuracy

Maximum average error: ±5mmHg

Maximum standard deviation: 8mmHg

Resolution: 1mmHg

Interval: 1, 2, 3, 4, 5, 10, 15, 30, 60, 90, 120, 180, 240, 480 minutes

Overpressure protection: Software and hardware, double safety protection

Cuff pressure range: 0-280mmHg

Northern SpO2

Measurement range: 0-100%

Resolution: 1%

Accuracy: ±2% (70-100%, Cat/Dog);

±3% (70-100%);

0-69%, unspecified

Refreshing Rate: 1s

Masimo SpO2 (optional)

Measurement range: 0-100%

Resolution: 1%

Accuracy: ±2% (70-100%, non-motion, low perfusion);

±3% (70-100%, non-motion);

±3% (70-100%, motion);

0-69%, unspecified

Refreshing Rate: 1s

Pulse Rate

Range: 30~254 bpm

Resolution: 1bpm

Accuracy: ±2bpm (non-motion)

±5bpm (motion)

Refreshing rate: 1s

Infrared Ear Thermometer (optional)

Displayed range: 34~42.2°C (93.2~108°F)

Operation ambient temperature range: 10~40°C (50~104°F)

Accuracy for displayed temperature range:

- ≥35°C (95.9°F) ~ ±42.2°C (107.6°F) range ±0.2°C (0.4°F)
- <35°C (95.9°F) ~ ±34°C (93.2°F) range ±0.3°C (0.5°F)

Phase II IRMA™ Sidestream CO2 (optional)

Warm-up time: Full accuracy within 10 seconds

Sampling flow rate: 50ml/min (+/-10/min)

Accuracy: 0~15% (±0.2% of the reading)

15~25%, unspecified

Measurement Range: 0-25%

Rise time: 200ms, typical at 50ml/min flow rate

Total response time:

within 3 seconds (with 2m Moline sampling line)

AWRR Range: 0-150bpm

Phase II IRMA™ Mainstream CO2 (optional)

Measurement Range: 0-25%

Warm-up time: Full accuracy within 10 seconds

Accuracy: 0~15% (±0.2% of the reading)

15~25%, unspecified

AWRR Range: 0-150bpm

Operation Environment

Power: AC 100-250V, 50/60Hz

Temperature: 5-40°C

Humidity: <85%



Northern Meditec Limited

THAI MEDICALS TRADING CO.,LTD
Address : 67/17 Moo.1 Mahasawat, Phuthamonthon
Nakhonpathom 73170 Thailand
Tel: +66 81-3451047
Web: www.tmt-med.com
Web: www.thaimedicaltrading.co.th
E-mail : infinity_bio@hotmail.com



- 120** Hours long trend
- 60** Mins short trend
- 1000** NIBP measurements
- 200** Alarm events

Northern Meditec Limited

